

### Certificate of Performance

## Phosphate 2 (Phos 2) Reagent

Product Code      KR107  
Lot Number            4067


Date            07.09.2004

Expiry Date      10.03.2005

Reagent	Lot Number
Phos 1	4065
Phos 2	4067
Test Method	LTP 05
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Abs.
0		0
0.5	P0.5 203 04	0.142
1	P1.0 203 04	0.282
1.5	P1.5 203 04	0.428
2	P2.0 203 04	0.569

### Calibration Curve



Test	Result	Specification	Pass / Fail
Intercept	0.000	-0.005 to 0.010	Pass
Sensitivity      Abs/ppm	0.277	0.260 to 0.320	Pass
Regression	1.000	0.990 to 1.001	Pass

*Tom Russell.*

Dr. T. Russell