

Certificate of Performance

Phosphate 2 (Phos 2) Reagent

Product Code **KR107**
 Lot Number **4102**


Date 03.11.2004

Expiry Date 03.05.2005

Reagent	Lot Number
Phos 1	4101
Phos 2	4102
Test Method	LTP 05
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Abs.
0		0.001
0.5	P0.5 279 04	0.149
1	P1.0 278 04	0.295
1.5	P1.5 279 04	0.443
2	P2.0 279 04	0.588

Calibration Curve



Test	Result	Specification	Pass / Fail
Intercept	0.002	-0.005 to 0.010	Pass
Sensitivity	Abs/ppm 0.294	0.260 to 0.320	Pass
Regression	1.000	0.990 to 1.001	Pass



Dr. T. Russell