

Sulphate(0 -1000mg/l)

Version: 2009

Test Definition

Test type	Photometric
Full name	Sulphate
Online name	Sulphate
Result unit	mg/l
Number of dec	3
Acceptance	Manual
Dilution 1+	0






Test in use	YES		
	low	High	Units
Test limit	0	10000	mg/l
Initial abs	0	2	A
Dilution limit	*	1000	mg/l
Secondary dil 1+	0	9	
Correction factor	1		
Correction bias	0		

Cal. Parameters

Cal. Type	Non-linear	Factor	
Repeat time(d)	0	Bias corr. In use	No
Point/Cal	Duplicate		
Acceptance	Manual	Abs error (mA)	*
		Rel error (%)	*
Type of Std	Series		
Std id		Response limit (mA)	
Concentration		Min	*
		Max	*

Test Flow

Blank	Yes
-------	-----

				
Sample	Blank	Reagent	Incubation	End pt.
Volume (µl)	Min abs	Sulphate	Time (sec)	λ 1 (nm)
12	*	Volume (µl)	300	520
Disp. with	Max abs	40		λ 2 (nm)
Water	*	Disp. with		none
Volume (µl)		Extra		Meas. type
108		Volume (µl)		Normal
Dil. With		20		
Water				
Extra wash				
No				