

Certificate of Performance

**Amm 1 Reagent**



Product Code 984640  
Lot Number 0017

Date 28.01.2010


Expiry Date 25.01.2011

Reagent	Lot Number
AMM 1	0017
Amm 2	0018
Test Method	LTP06
Test Instrument	Konelab 20
Serial Number	P4919028

Std. Value	Lot #	Abs.
0.00		0.016
0.25	AMM 10 10 09	0.220
0.50	AMM 10 10 09	0.428
0.77	AMM 10 10 09	0.653
1.00	AMM 10 10 09	0.843

\*serial dilution

Calibration Curve



Test	Result	Specification	Pass / Fail
Intercept	0.0146	0.000 to 0.025	Pass
Sensitivity Abs/ppm	0.8282	0.750 to 0.890	Pass
Regression	1.0000	0.990 to 1.0001	Pass

Notes Store between 2 & 8 °c  
\*Stable on Board 1 day  
Barcode ID = 751  
\* Based on results obtained in our Research Facility

*Tom Russell*  
Dr. T. Russell